

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	1.00006	-0.060	± 1	±	4.3E-01 mm						
0 a 2	2.00064	-0.640	± 1.4	±	4.8E-01 mm						
0 a 3	2.99992	0.080	± 1.8	±	4.8E-01 mm						
0 a 4	4.00037	-0.370	± 2.2	±	4.8E-01 mm						
0 a 5	4.99987	0.130	± 2.6	±	4.8E-01 mm						
0 a 6	6.00050	-0.500	± 3	±	3.6E-01 mm						
0 a 7	7.00043	-0.430	± 3.4	±	3.6E-01 mm						
0 a 8	8.00089	-0.890	± 3.8	±	3.6E-01 mm						
0 a 9	9.00003	-0.030	± 4.2	±	3.6E-01 mm						
0 a 10	10.00040	-0.400	± 4.6	±	3.6E-01 mm						
0 a 11	11.00057	-0.260	± 5	±	3.6E-01 mm						
0 a 12	12.00099	-0.720	± 5.4	±	3.6E-01 mm						
0 a 13	13.00054	-0.310	± 5.8	±	3.6E-01 mm						
0 a 14	14.00123	-1.050	± 6.2	±	3.6E-01 mm						
0 a 15	15.00099	-0.840	± 6.6	±	3.6E-01 mm						