

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	24.0 μm	0.7 μm	± 1.23 μm	±	1.6E+00 μm
51.6 μm	51.7 μm	0.1 μm	± 2.52 μm	±	6.2E-01 μm
124.6 μm	125.6 μm	1.0 μm	± 3.25 μm	±	1.3E+00 μm
250.3 μm	250.3 μm	0.0 μm	± 4.5 μm	±	6.2E-01 μm
486 μm	481.5 μm	-4.5 μm	± 6.86 μm	±	3.5E+00 μm
963 μm	956.8 μm	-6.3 μm	± 11.63 μm	±	4.1E+00 μm