

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 23.3 μm           | 24.4 μm   | 1.1 μm        | ± 0.73 μm         | ±                           | 5.8E-02 μm |
| 51.6 μm           | 51.4 μm   | -0.2 μm       | ± 1.02 μm         | ±                           | 5.8E-02 μm |
| 124.6 μm          | 124.8 μm  | 0.2 μm        | ± 5.7 μm          | ±                           | 6.1E-02 μm |
| 250.3 μm          | 250.7 μm  | 0.4 μm        | ± 9.5 μm          | ±                           | 6.9E-02 μm |
| 486 μm            | 485.4 μm  | -0.6 μm       | ± 16.6 μm         | ±                           | 9.4E-02 μm |
| 952 μm            | 955.0 μm  | 3.0 μm        | ± 30.6 μm         | ±                           | 1.6E-01 μm |