

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.853 °C	-0.7 °C	0.153 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.522 °C	15.2 °C	-0.322 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	29.9 °C	-0.463 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	32.1 %HR	1.939 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.376 %HR	51.7 %HR	1.324 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.754 %HR	72.0 %HR	1.246 %HR	± 3.0 %HR	± 1.9E+00 %HR