

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.8 °C	-0.262 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.0 %HR	1.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.345 %HR	50.9 %HR	0.555 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	68.2 %HR	-2.145 %HR	± 3.0 %HR	± 1.9E+00 %HR