

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 193.2 lx                 | 200 lx           | -6.8 lx              | ± 10 lx                     | ±                                  | 5.8E-01 lx |
| 340.7 lx                 | 345 lx           | -4.3 lx              | ± 16 lx                     | ±                                  | 6.0E-01 lx |
| 754.5 lx                 | 760 lx           | -5.5 lx              | ± 32 lx                     | ±                                  | 9.5E-01 lx |
| 926.6 lx                 | 940 lx           | -13.4 lx             | ± 39 lx                     | ±                                  | 1.1E+00 lx |
| 2,878.7 lx               | 2,900 lx         | -21.3 lx             | ± 125 lx                    | ±                                  | 7.6E+00 lx |
| 51,091.6 lx              | 50,500 lx        | 591.6 lx             | ± 2144 lx                   | ±                                  | 4.4E+02 lx |