

| Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | E.M.P ( <i>Tolerance</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |            |
|----------------------------|--------------------|------------------------|----------------------------|--------------------------------------|------------|
| 23.3 μm                    | 23.2 μm            | -0.1 μm                | ± 2.5 μm                   | ±                                    | 5.8E-02 μm |
| 51.6 μm                    | 51.8 μm            | 0.2 μm                 | ± 2.5 μm                   | ±                                    | 5.8E-02 μm |
| 124.6 μm                   | 125 μm             | 0.4 μm                 | ± 3.7 μm                   | ±                                    | 6.1E-02 μm |
| 250.3 μm                   | 250 μm             | -0.3 μm                | ± 7.5 μm                   | ±                                    | 6.9E-02 μm |
| 486.0 μm                   | 487 μm             | 1.0 μm                 | ± 15 μm                    | ±                                    | 9.4E-02 μm |
| 963.0 μm                   | 964 μm             | 1.0 μm                 | ± 29 μm                    | ±                                    | 1.6E-01 μm |