

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0098 in	0.0098 in	0.0000 in	± 0.0005 in	±	5.8E-05 in
0.0188 in	0.0188 in	0.0000 in	± 0.0009 in	±	5.8E-05 in
0.0375 in	0.0375 in	0.0000 in	± 0.0019 in	±	5.8E-05 in
0.0625 in	0.0626 in	0.0001 in	± 0.0031 in	±	5.8E-05 in
0.1000 in	0.1002 in	0.0002 in	± 0.0050 in	±	5.8E-05 in
0.1250 in	0.1251 in	0.0001 in	± 0.0063 in	±	7.6E-02 in
0.1500 in	0.1501 in	0.0001 in	± 0.0075 in	±	1.0E-01 in
0.2000 in	0.2002 in	0.0002 in	± 0.0100 in	±	1.0E-01 in
0.2500 in	0.2502 in	0.0002 in	± 0.0125 in	±	9.4E-02 in