

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.54 μm	-0.11 μm	± 2.5 μm	±	5.8E-01 μm
155.23 μm	154.6 μm	-0.63 μm	± 2.5 μm	±	5.8E-01 μm
257.12 μm	257.8 μm	0.68 μm	± 2.5 μm	±	5.8E-01 μm
511.72 μm	512.0 μm	0.28 μm	± 2.5 μm	±	5.8E-01 μm