

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.5 °C	-0.210 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.601 °C	0.4 °C	-0.201 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.545 °C	5.4 °C	-0.145 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.037 °C	10.3 °C	0.263 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.400 °C	25.1 °C	-0.300 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.9 °C	0.336 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.591 °C	40.9 °C	0.309 °C	± 0.45 °C	± 4.0E-01 °C

