

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.5 °C	-0.210 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.601 °C	0.3 °C	-0.301 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.545 °C	5.2 °C	-0.345 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.037 °C	9.8 °C	-0.237 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.400 °C	25.5 °C	0.100 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.7 °C	0.337 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.591 °C	40.2 °C	-0.391 °C	± 0.45 °C	± 4.0E-01 °C

