

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-26.0 °C	-0.710 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.601 °C	0.2 °C	-0.401 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.545 °C	5.9 °C	0.355 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.037 °C	9.7 °C	-0.337 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.400 °C	25.5 °C	0.100 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.8 °C	0.236 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.591 °C	40.3 °C	-0.291 °C	± 0.45 °C	± 4.0E-01 °C

