

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.9 °C	-0.610 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.662 °C	-0.2 °C	0.462 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.0 °C	0.007 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.037 °C	10.1 °C	0.063 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.2 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.032 °C	30.4 °C	0.368 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.140 °C	40.6 °C	0.460 °C	± 0.5 °C	± 4.0E-01 °C

