

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-24.7 °C	0.590 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.662 °C	-0.4 °C	0.262 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.0 °C	0.007 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.037 °C	9.8 °C	-0.237 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.6 °C	0.351 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.032 °C	30.5 °C	0.468 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.140 °C	40.1 °C	-0.040 °C	± 0.5 °C	± 4.0E-01 °C

