

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.290 °C        | -25.2 °C  | 0.090 °C      | ± 0.70 °C            | ± 4.0E-01 °C                |
| 0 °C              | -0.662 °C         | -0.7 °C   | -0.038 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 5.3 °C    | 0.307 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.037 °C         | 10.4 °C   | 0.363 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.249 °C         | 25.5 °C   | 0.251 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.032 °C         | 30.5 °C   | 0.468 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.140 °C         | 40.6 °C   | 0.460 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

