

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.3 °C	-0.010 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.662 °C	-0.7 °C	-0.038 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.0 °C	0.007 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.037 °C	10.2 °C	0.163 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.3 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.032 °C	29.9 °C	-0.132 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.140 °C	40.1 °C	-0.040 °C	± 0.5 °C	± 4.0E-01 °C

