

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.587 °C	3.5 °C	0.913 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.509 °C	11.3 °C	0.791 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.754 %HR	22.0 %HR	1.246 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	75.0 %HR	-5.673 %HR	± 5.0 %HR	± 1.9E+00 %HR