

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.522 °C	16.5 °C	0.978 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.6 °C	0.036 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.754 %HR	20.0 %HR	-0.754 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	64.313 %HR	60.7 %HR	-3.613 %HR	± 5.0 %HR	± 1.9E+00 %HR