

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.522 °C	15.1 °C	-0.422 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.564 °C	29.7 °C	-0.864 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.754 %HR	21.0 %HR	0.246 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	89.0 %HR	8.327 %HR	± 5.0 %HR	± 1.9E+00 %HR