

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	14.8 °C	-0.722 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.564 °C	29.7 °C	-0.864 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.754 %HR	26.0 %HR	5.246 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	87.0 %HR	6.327 %HR	± 5.0 %HR	± 1.9E+00 %HR