

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	13.6 °C	-1.922 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.564 °C	28.9 °C	-1.664 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.754 %HR	21.0 %HR	0.246 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	86.0 %HR	5.327 %HR	± 5.0 %HR	± 1.9E+00 %HR