

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	9.997 °C	9.7 °C	-0.297 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	19.7 °C	-0.325 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.293 °C	30.1 °C	-0.193 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.280 °C	39.8 °C	-0.480 °C	± 0.5 °C	± 4.0E-01 °C