

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	9.977 °C	10.4 °C	0.423 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.396 °C	20.4 °C	0.004 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.433 °C	30.5 °C	0.067 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.381 °C	40.1 °C	-0.281 °C	± 0.5 °C	± 4.0E-01 °C