

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.977 °C	10.4 °C	0.423 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.396 °C	20.6 °C	0.204 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.433 °C	30.8 °C	0.367 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.381 °C	40.3 °C	-0.081 °C	± 0.5 °C	± 4.0E-01 °C