

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	30.1 °C	-0.193 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	25.856 %HR	28 %HR	2.144 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.556 %HR	50 %HR	1.444 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.726 %HR	76 %HR	4.274 %HR	± 5.0 %HR	± 1.9E+00 %HR