

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.363 °C	29.9 °C	-0.463 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	30.3 %HR	0.139 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.376 %HR	49.0 %HR	-1.376 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	63.0 %HR	-2.335 %HR	± 5.0 %HR	± 1.9E+00 %HR