

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09920	0.800	± 1.2	± 4.3E-01 mm					
0 a 0.5	0.49897	1.030	± 1.6	± 4.3E-01 mm					
0 a 1	0.99907	0.930	± 2	± 4.3E-01 mm					
0 a 2	1.99906	0.940	± 2.8	± 4.8E-01 mm					
0 a 3	2.99915	0.850	± 3.6	± 4.8E-01 mm					
0 a 4	3.99933	0.670	± 4.4	± 4.8E-01 mm					
0 a 5	4.99923	0.770	± 5.2	± 4.8E-01 mm					
0 a 6	5.99923	0.770	± 6	± 3.6E-01 mm					
0 a 7	6.99942	0.580	± 6.8	± 3.6E-01 mm					
0 a 8	0.62000	0.620	± 7.6	± 3.6E-01 mm					