

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49932	0.680	± 1.6	± 4.3E-01 mm					
0 a 1	0.99897	1.030	± 2	± 4.3E-01 mm					
0 a 1.5	1.49919	0.810	± 2.4	± 4.8E-01 mm					
0 a 2	1.99891	1.090	± 2.8	± 4.8E-01 mm					
0 a 2.5	2.49941	0.590	± 3.2	± 4.8E-01 mm					
0 a 3	2.99915	0.850	± 3.6	± 4.8E-01 mm					
0 a 3.5	3.49962	0.380	± 4	± 4.8E-01 mm					
0 a 4	3.99903	0.970	± 4.4	± 4.8E-01 mm					
0 a 4.5	4.49951	0.490	± 4.8	± 4.8E-01 mm					
0 a 5	4.99913	0.870	± 5.2	± 4.8E-01 mm					