

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	14.8 °C	-0.722 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.1 °C	-0.263 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	29.0 %HR	-1.161 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.642 %HR	62.0 %HR	-3.642 %HR	± 5.0 %HR	± 1.9E+00 %HR
85 %HR	80.366 %HR	80.0 %HR	-0.366 %HR	± 5.0 %HR	± 1.9E+00 %HR