

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.60 μm	49.60 μm	-1.00 μm	± 2.5 μm	±	5.8E-01 μm
144.40 μm	144.6 μm	0.20 μm	± 2.5 μm	±	5.8E-01 μm
503.10 μm	246.8 μm	-256.30 μm	± 2.5 μm	±	5.8E-01 μm
725.20 μm	724.8 μm	-0.40 μm	± 2.5 μm	±	5.9E-01 μm