

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50.60 μm	49.60 μm	-1.00 μm	± 2.5 μm	±	5.8E-01 μm
144.40 μm	144.6 μm	0.20 μm	± 2.5 μm	±	5.8E-01 μm
503.10 μm	502.4 μm	-0.70 μm	± 2.5 μm	±	5.8E-01 μm
725.20 μm	724.8 μm	-0.40 μm	± 2.5 μm	±	5.9E-01 μm