

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50.60 μm	50.50 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
144.40 μm	145.0 μm	0.60 μm	± 2.5 μm	±	5.8E-01 μm
246.20 μm	246.8 μm	0.60 μm	± 2.5 μm	±	5.8E-01 μm
503.10 μm	503.2 μm	0.10 μm	± 2.5 μm	±	5.8E-01 μm