

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.951 °C	16.0 °C	1.049 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.788 °C	25.3 °C	0.512 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.937 °C	49.9 °C	-0.037 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.744 %HR	24 %HR	-1.744 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.098 %HR	46 %HR	-3.098 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.824 %HR	70 %HR	-4.824 %HR	± 5.0 %HR	± 1.9E+00 %HR