

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.951 °C	15.9 °C	0.949 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.788 °C	25.4 °C	0.612 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.937 °C	50.5 °C	0.563 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.744 %HR	24 %HR	-1.744 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.098 %HR	47 %HR	-2.098 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.824 %HR	71 %HR	-3.824 %HR	± 5.0 %HR	± 1.9E+00 %HR