

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.951 °C	16.0 °C	1.049 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.788 °C	25.7 °C	0.912 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.937 °C	50.9 °C	0.963 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.744 %HR	26 %HR	0.256 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.098 %HR	46 %HR	-3.098 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.824 %HR	72 %HR	-2.824 %HR	± 5.0 %HR	± 1.9E+00 %HR