

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.017	0.01628	0.720	± 1.2	± 4.3E-01 mm					
0 a 0.034	0.03330	0.700	± 1.2	± 4.3E-01 mm					
0 a 0.051	0.05034	0.660	± 1.2	± 4.3E-01 mm					
0 a 0.068	0.06735	0.650	± 1.2	± 4.3E-01 mm					
0 a 0.085	0.08441	0.590	± 1.2	± 4.3E-01 mm					
0 a 0.102	0.10143	0.570	± 1.2	± 4.3E-01 mm					
0 a 0.119	0.11845	0.550	± 1.2	± 4.3E-01 mm					
0 a 0.136	0.13551	0.490	± 1.3	± 4.3E-01 mm					
0 a 0.153	0.15256	0.440	± 1.3	± 4.3E-01 mm					
0 a 0.17	0.16962	0.380	± 1.3	± 4.3E-01 mm					