

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.790 °C	15.0 °C	0.210 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.878 °C	25.0 °C	0.122 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.887 °C	31.9 °C	0.013 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	19.578 %HR	22.5 %HR	2.922 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.090 %HR	51.4 %HR	1.310 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.908 %HR	68.0 %HR	-2.908 %HR	± 5.0 %HR	± 1.9E+00 %HR