

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	30.714 °C	31.0 °C	0.286 °C	±	1 °C	± 4.0E-01 °C
35 %HR	36.511 %HR	39.2 %HR	2.689 %HR	±	3 %HR	± 1.9E+00 %HR
50 %HR	52.799 %HR	55.1 %HR	2.301 %HR	±	3 %HR	± 1.9E+00 %HR
70 %HR	73.433 %HR	71.4 %HR	-2.033 %HR	±	3 %HR	± 1.9E+00 %HR