

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>		Incertidumbre <i>(Uncertainty)</i>
30 °C	30.714 °C	30.6 °C	-0.114 °C	±	1 °C	± 4.0E-01 °C
35 %HR	36.511 %HR	39.9 %HR	3.389 %HR	±	3 %HR	± 1.9E+00 %HR
50 %HR	52.799 %HR	55.9 %HR	3.101 %HR	±	3 %HR	± 1.9E+00 %HR
70 %HR	73.433 %HR	74.7 %HR	1.267 %HR	±	3 %HR	± 1.9E+00 %HR