

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.8	0.79920	1.070	± 1.8	±	#REF! mm						
0 a 0.9	0.89910	1.180	± 1.9	±	#REF! mm						
0 a 1	0.99900	1.280	± 2	±	#REF! mm						
0 a 1.1	1.09890	1.380	± 2	±	#REF! mm						
0 a 1.2	1.19880	1.490	± 2.1	±	#REF! mm						
0 a 1.3	1.29870	1.590	± 2.2	±	#REF! mm						
0 a 1.4	1.39860	1.700	± 2.3	±	#REF! mm						
0 a 1.5	1.49850	1.800	± 2.4	±	#REF! mm						
0 a 1.6	1.59840	1.910	± 2.4	±	#REF! mm						
0 a 1.7	1.69830	2.010	± 2.5	±	#REF! mm						
0 a 1.8	1.79820	2.120	± 2.6	±	#REF! mm						
0 a 1.9	1.89810	2.230	± 2.7	±	#REF! mm						
0 a 2	1.99800	2.330	± 2.8	±	#REF! mm						
0 a 2.1	2.09790	2.440	± 2.8	±	#REF! mm						