

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.290 °C        | -25.6 °C  | -0.310 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.221 °C         | -0.5 °C   | -0.279 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 5.0 °C    | 0.007 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.007 °C         | 10.2 °C   | 0.193 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.349 °C         | 25.1 °C   | -0.249 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.062 °C         | 30.3 °C   | 0.238 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.291 °C         | 40.1 °C   | -0.191 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

