

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.333 °C	30.6 °C	0.267 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.210 %HR	30.0 %HR	0.790 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.640 %HR	50.6 %HR	0.960 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.263 %HR	65.4 %HR	0.137 %HR	± 5.0 %HR	± 1.9E+00 %HR