

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.333 °C	30.8 °C	0.467 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.210 %HR	28.5 %HR	-0.710 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.640 %HR	49.6 %HR	-0.040 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.263 %HR	64.7 %HR	-0.563 %HR	± 5.0 %HR	± 1.9E+00 %HR