

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	34.745 °C	35.4 °C	0.655 °C	±	1 °C	± 4.0E-01 °C
35 %HR	36.204 %HR	37.0 %HR	0.796 %HR	±	5 %HR	± 1.9E+00 %HR
50 %HR	52.094 %HR	50.0 %HR	-2.094 %HR	±	5 %HR	± 1.9E+00 %HR
70 %HR	73.433 %HR	75.0 %HR	1.567 %HR	±	5 %HR	± 1.9E+00 %HR