

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) |       | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|-------|--------------------------------------|
| 30 °C                      | 34.745 °C                  | 35.4 °C            | 0.655 °C               | ±                             | 1 °C  | ± 4.0E-01 °C                         |
| 30 %HR                     | 36.204 %HR                 | 37.0 %HR           | 0.796 %HR              | ±                             | 5 %HR | ± 1.9E+00 %HR                        |
| 50 %HR                     | 52.094 %HR                 | 50.0 %HR           | -2.094 %HR             | ±                             | 5 %HR | ± 1.9E+00 %HR                        |
| 70 %HR                     | 73.433 %HR                 | 75.0 %HR           | 1.567 %HR              | ±                             | 5 %HR | ± 1.9E+00 %HR                        |