

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	34.745 °C	35.2 °C	0.455 °C	± 1 °C	± 4.0E-01 °C
30 %HR	36.204 %HR	33.0 %HR	-3.204 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	52.094 %HR	48.0 %HR	-4.094 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	73.433 %HR	70.0 %HR	-3.433 %HR	± 5 %HR	± 1.9E+00 %HR