

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.017	0.01633	0.670	± 1.2	± 4.3E-01 mm					
0 a 0.034	0.03347	0.530	± 1.2	± 4.3E-01 mm					
0 a 0.051	0.05052	0.480	± 1.2	± 4.3E-01 mm					
0 a 0.068	0.06770	0.300	± 1.2	± 4.3E-01 mm					
0 a 0.085	0.08439	0.610	± 1.2	± 4.3E-01 mm					
0 a 0.102	0.10141	0.590	± 1.2	± 4.3E-01 mm					
0 a 0.119	0.11853	0.470	± 1.2	± 4.3E-01 mm					
0 a 0.136	0.13548	0.520	± 1.3	± 4.3E-01 mm					
0 a 0.153	0.15258	0.420	± 1.3	± 4.3E-01 mm					
0 a 0.17	0.16959	0.410	± 1.3	± 4.3E-01 mm					