

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.010 °C	-0.1 °C	-0.110 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	5.0 °C	-0.144 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.188 °C	10.2 °C	0.012 °C	± 0.5 °C	± 4.0E-01 °C