

3/28/2025

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.010 °C	0.1 °C	0.090 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	4.9 °C	-0.244 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.188 °C	9.9 °C	-0.288 °C	± 0.5 °C	± 4.0E-01 °C