

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) |       | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|-------|--------------------------------------|
| 30 °C                      | 30.714 °C                  | 30.7 °C            | -0.014 °C              | ±                             | 1 °C  | ± 4.0E-01 °C                         |
| 30 %HR                     | 36.511 %HR                 | 38.6 %HR           | 2.089 %HR              | ±                             | 3 %HR | ± 1.9E+00 %HR                        |
| 50 %HR                     | 52.799 %HR                 | 51.0 %HR           | -1.799 %HR             | ±                             | 3 %HR | ± 1.9E+00 %HR                        |
| 70 %HR                     | 73.433 %HR                 | 71.3 %HR           | -2.133 %HR             | ±                             | 3 %HR | ± 1.9E+00 %HR                        |